



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Accura Calibration**  
**2834 West Kingsley Road**  
**Garland, TX 75041**

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standards

**ANSI/NCSL Z540-1-1994 (R2002) and**  
**ANSI/NCSL Z540.3-2006 (R2013)**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 01 February 2024  
Certificate Number: AC-2548



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**  
**AND**  
**ANSI/NCSL Z540-1-1994 (R2002)**  
**ANSI/NCSL Z540.3-2006 (R2013)**

**Accura Calibration**  
 2834 West Kingsley Road  
 Garland, TX 75041  
 Dwight Martin 972-278-7878

**CALIBRATION**

Valid to: **February 1, 2024**

Certificate Number: **AC-2548**

**Acoustics and Vibration**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers Frequency Response (1 g Reference)	(7 to 10) Hz (10 to 30) Hz (30 to 99) Hz 100 Hz (101 to 2 000) Hz (2 000 to 10 000) Hz	7 % of reading 4.6 % of reading 3.8 % of reading 3.8 % of reading 3.8 % of reading 5.3 % of reading	Modal Shop 9110D Workstation

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Clamp Meter <sup>1</sup> Toroidal Clamp (45 to 65) Hz	(20 to 300) A (300 to 600) A (600 to 1 000) A	11 mA/A + 90 mA 9.8 mA/A + 90 mA 11 mA/A + 0.09 A	Fluke 5500A/Coil, Valhalla 2575A Current Shunt, Keysight 3458A 8.5 Digit Multimeter, Current Source
400 Hz	(20 to 300) A (300 to 600) A	19 mA/A + 0.1 A 18 mA/A + 0.1 A	
200 Hz	(600 to 1 000) A	18 mA/A + 0.1 A	



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Clamp Meter <sup>1</sup> Non-Toroidal Clamp (45 to 65) Hz	(20 to 300) A (300 to 600) A (600 to 1 000) A	15 mA/A + 0.9 A 14 mA/A + 0.9 A 14 mA/A + 0.9 A	Fluke 5500A/Coil, Valhalla 2575A Current Shunt, Keysight 3458A 8.5 Digit Multimeter, Current Source
400 Hz	(20 to 300) A (300 to 600) A	22 mA/A + 0.9 A 21 mA/A + 0.9 A	
200Hz	(600 to 1 000) A	22 mA/A + 0.9 A	
DC Voltage – Source <sup>1</sup> (Locked Ranges)	(0 to 329.999 9) mV (0 to 3.299 999) V (0 to 32.999 99) V (30 to 329.999 9) V (100 to 1 020) V	16 μV/V + 0.78 μV 9.1 μV/V + 1.6 μV 9.7 μV/V + 16 μV 14 μV/V + 0.12 μV 15 μV/V + 1.2 μV	Fluke Multiproduct Calibrator
DC Voltage – Measure <sup>1</sup>	Up to 120 mV (0.12 to 1.2) V (1.2 to 12) V (12 to 120) V (120 to 1 050) V	6.1 μV/V + 0.3 μV 5.1 μV/V + 0.3 μV 5.1 μV/V + 0.5 μV 7.3 μV/V + 30 μV 7.3 μV/V + 0.1 mV	Keysight 3458A 8.5 Digit Multimeter
DC High Voltage – Source/Measure <sup>1</sup>	(1 to 60) kV	1.8 mV/V	Ross Engineering VD60-6.2Y-A-LB-AL High Voltage Divider, Keysight 3458A 8.5 Digit Multimeter
DC Current – Source <sup>1</sup> (Locked Ranges)	(0 to 329) μA (0 to 3.299 999) mA (0 to 32.999 99) mA (0 to 329.999 9) mA (0 to 1.099 999) A (1.1 to 2.999 99) A (0 to 10.999 9) A (11 to 20.5) A	0.12 μA/μA + 16 nA 79 μA/μA + 39 nA 79 μA/μA + 0.19 μA 82 μA/μA + 1.9 μA 0.15 mA/A + 31 μA 0.3 mA/A + 31 μA 0.39 mA/A + 0.39 mA 0.78 mA/A + 0.58 mA	Fluke Multiproduct Calibrator
DC Current – Measure <sup>1</sup>	(0.1 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A (2 to 20) A (20 to 100) A	3.9 mA/A + 3 nA 0.41 mA/A + 30 nA 0.11 mA/A + 0.3 μA 0.2 mA/A + 3 μA 0.2 mA/A + 30 μA 0.5 mA/A + 0.3 mA	Keysight 3458A 8.5 Digit Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(1 to 32.999) mV		Fluke Multiproduct Calibrator
	(10 to 45) Hz	0.64 mV/V + 4.7 μV	
	45 Hz to 10 kHz	0.15 mV/V + 4.7 μV	
	(10 to 20) kHz	0.18 mV/V + 4.7 μV	
	(20 to 50) kHz	0.79 mV/V + 4.7 μV	
	(50 to 100) kHz	2.7 mV/V + 9.3 μV	
	(100 to 500) kHz	6.4 mV/V + 39 μV	
	(33 to 329.999) mV		
	(10 to 45) Hz	0.36 mV/V + 6.2 μV	
	45 Hz to 10 kHz	0.3 mV/V + 6.2 μV	
	(10 to 20) kHz	0.3 mV/V + 6.2 μV	
	(20 to 50) kHz	0.52 mV/V + 6.2 μV	
	(50 to 100) kHz	0.97 mV/V + 27 μV	
	(100 to 500) kHz	2.8 mV/V + 54 μV	
	(0.33 to 3.299 99) V		
	(10 to 45) Hz	0.23 mV/V + 39 μV	
	45 Hz to 10 kHz	0.12 mV/V + 47 μV	
	(10 to 20) kHz	0.15 mV/V + 47 μV	
	(20 to 50) kHz	0.24 mV/V + 39 μV	
	(50 to 100) kHz	0.55 mV/V + 97 μV	
	(100 to 500) kHz	1.9 mV/V + 0.47 mV	
	(3.3 to 32.999 9) V		
	(10 to 45) Hz	0.23 mV/V + 0.5 mV	
	45 Hz to 10 kHz	0.12 mV/V + 0.47 mV	
(10 to 20) kHz	0.2 mV/V + 4.7 mV		
(20 to 50) kHz	0.23 mV/V + 4.7 mV		
(50 to 100) kHz	0.7 mV/V + 1.2 mV		
(33 to 329.999) V			
45 Hz to 1 kHz	0.15 mV/V + 1.6 mV		
(1 to 10) kHz	0.16 mV/V + 4.7 mV		
(10 to 20) kHz	0.2 mV/V + 4.7 mV		
(20 to 50) kHz	0.25 mV/V + 4.7 mV		
(50 to 100) kHz	0.7 mV/V + 39 mV		
(330 to 1 020) V			
45 Hz to 1 kHz	0.24 mV/V + 7.8 mV		
(1 to 5) kHz	0.2 mV/V + 7.8 mV		
(5 to 10) kHz	0.24 mV/V + 7.8 mV		



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	10 nV to 12 mV		Keysight 3458A 8.5 Digit Multimeter
	(1 to 40) Hz	0.7 mV/V	
	40 Hz to 1 kHz	0.36 mV/V	
	(1 to 20) kHz	0.48 mV/V	
	(20 to 50) kHz	1.3 mV/V	
	(50 to 100) kHz	5.9 mV/V	
	(100 to 300) kHz	46 mV/V	
	12 mV to 12 V		
	(1 to 40) Hz	0.13 mV/V	
	40 Hz to 1 kHz	0.11 mV/V	
	(1 to 20) kHz	0.22 mV/V	
	(20 to 50) kHz	0.37 mV/V	
	(50 to 100) kHz	0.95 mV/V	
	(100 to 300) kHz	3.6 mV/V	
	(0.3 to 1) MHz	1.2 mV/V	
	(1 to 2) MHz	17 mV/V	
	(12 to 120) V		
	(1 to 40) Hz	0.28 mV/V	
40 Hz to 1 kHz	0.26 mV/V		
(1 to 20) kHz	0.26 mV/V		
(20 to 50) kHz	0.43 mV/V		
(50 to 100) kHz	1.4 mV/V		
(100 to 300) kHz	4.7 mV/V		
(0.3 to 1) MHz	17 mV/V		
(120 to 700) V			
(1 to 40) Hz	0.53 mV/V		
40 Hz to 1 kHz	0.5 mV/V		
(1 to 20) kHz	0.73 mV/V		
(20 to 50) kHz	1.4 mV/V		
(50 to 100) kHz	3.5 mV/V		
AC High Voltage – Source/Measure <sup>1</sup>	(1 to 60) kV 60 Hz	5.9 mV/V	Ross Engineering VD60-6.2Y-A-LB-AL High Voltage Divider, Keysight 3458A 8.5 Digit Multimeter

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1</sup>	(29 to 329.99) $\mu$ A		Fluke Multiproduct Calibrator
	(10 to 20) Hz	1.6 nA/ $\mu$ A + 78 nA	
	(20 to 45) Hz	1.2 nA/ $\mu$ A + 78 nA	
	45 Hz to 1 kHz	0.97 nA/ $\mu$ A + 78 nA	
	(1 to 5) kHz	2.3 nA/ $\mu$ A + 0.12 $\mu$ A	
	(5 to 10) kHz	6.4 nA/ $\mu$ A + 0.16 $\mu$ A	
	(10 to 30) kHz	12 nA/ $\mu$ A + 0.31 $\mu$ A	
	(0.33 to 3.299 99) mA		
	(10 to 20) Hz	1.6 $\mu$ A/mA + 0.12 $\mu$ A	
	(20 to 45) Hz	0.97 $\mu$ A/mA + 0.12 $\mu$ A	
	45 Hz to 1 kHz	0.79 $\mu$ A/mA + 78 nA	
	(1 to 5) kHz	1.5 $\mu$ A/mA + 0.16 $\mu$ A	
	(5 to 10) kHz	3.9 $\mu$ A/mA + 0.23 $\mu$ A	
	(10 to 30) kHz	7.9 $\mu$ A/mA + 0.47 $\mu$ A	
	(3.3 to 32.999 9) mA		
	(10 to 20) Hz	1.4 $\mu$ A/mA + 1.6 $\mu$ A	
	(20 to 45) Hz	0.7 $\mu$ A/mA + 1.6 $\mu$ A	
	45 Hz to 1 kHz	0.3 $\mu$ A/mA + 0.16 $\mu$ A	
	(1 to 5) kHz	0.64 $\mu$ A/mA + 0.16 $\mu$ A	
	(5 to 10) kHz	1.5 $\mu$ A/mA + 2.3 $\mu$ A	
	(10 to 30) kHz	3 $\mu$ A/mA + 3.1 $\mu$ A	
	(33 to 329.999) mA		
	(10 to 20) Hz	1.4 $\mu$ A/mA + 16 $\mu$ A	
	(20 to 45) Hz	0.7 $\mu$ A/mA + 16 $\mu$ A	
45 Hz to 1 kHz	0.33 $\mu$ A/mA + 16 $\mu$ A		
(1 to 5) kHz	0.79 $\mu$ A/mA + 39 $\mu$ A		
(5 to 10) kHz	1.6 $\mu$ A/mA + 78 $\mu$ A		
(10 to 30) kHz	3.3 $\mu$ A/mA + 0.16 mA		
(0.33 to 1.099 99) A			
(10 to 45) Hz	0.11 mA/A + 78 $\mu$ A		
45 Hz to 1 kHz	0.11 mA/A + 78 $\mu$ A		
(1 to 5) kHz	0.16 mA/A + 0.78 mA		
(5 to 10) kHz	0.4 mA/A + 3.9 mA		
(1.1 to 2.999 99) A			
(10 to 45) Hz	1.4 mA/A + 78 $\mu$ A		
45 Hz to 1 kHz	0.53 mA/A + 78 $\mu$ A		
(1 to 5) kHz	4.7 mA/A + 0.78 mA		
(5 to 10) kHz	19 mA/A + 3.9 mA		



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1</sup>	(3 to 10.9999) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (3 to 10.9999) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (11 to 20.5) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz	1 mA/A + 1.6 mA 1.2 mA/A + 1.6 mA 23 mA/A + 1.6 mA  1 mA/A + 1.6 mA 1.2 mA/A + 1.6 mA 23 mA/A + 1.6 mA  1.8 mA/A + 3.9 mA 1.9 mA/A + 3.9 mA 23 mA/A + 3.9 mA	Fluke Multiproduct Calibrator
AC Current – Source/Measure <sup>1</sup>	45 Hz to 1 kHz (0.1 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A (2 to 20) A (20 to 100) A	4.1 mA/A + 3 nA 1.1 mA/A + 30 nA 1 mA/A + 0.3 μA 1 mA/A + 3 μA 1 mA/A + 30 μA 1 mA/A + 0.15 mA	Valhalla 2575A Current Shunt, Keysight 3458A 8.5 Digit Multimeter, Current Source
DC Power – Source <sup>1</sup> 33 mV to 1 020 V	0.33 mA to 2.9999 A 11 μW to 3.06 kW	0.018 % of reading	Fluke Multiproduct Calibrator
33 mV to 1 020 V	(3 to 20.5) A 0.1 W to 20.9 kW	0.054 % of reading	
AC Power – Source <sup>1</sup> 33 mV to 1 020 V (45 to 65) Hz Power Factor: PF=1	(3 to 20.5) A 0.11 mW to 20.9 kW	0.11 % of reading	Fluke Multiproduct Calibrator





ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source <sup>1</sup> 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 3 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz (10 to 600) Hz (10 to 300) Hz (10 to 150) Hz (10 to 120) Hz (10 to 80) Hz (0 to 50) Hz (0 to 20) Hz (0 to 6) Hz (0 to 2) Hz (0 to 0.6) Hz (0 to 0.2) Hz	(220 to 399.9) pF (0.4 to 1.099 9) nF (1.1 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) $\mu$ F (1.1 to 3.299 99) $\mu$ F (3.3 to 10.999 9) $\mu$ F (11 to 32.999 9) $\mu$ F (33 to 109.999) $\mu$ F (110 to 329.999) $\mu$ F (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.55 % of reading + 7.8 pF 0.42 % of reading + 10 pF 0.39 % of reading + 10 pF 0.2 % of reading + 10 pF 0.2 % of reading + 8 pF 0.2 % of reading + 8 pF 0.27 % of reading + 0.23 nF 0.21 % of reading + 0.78 nF 0.2 % of reading + 2.3 nF 0.21 % of reading + 7.8 nF 0.3 % of reading + 23 nF 0.35 % of reading + 78 nF 0.36 % of reading + 0.23 $\mu$ F 0.36 % of reading + 0.78 $\mu$ F 0.36 % of reading + 2.3 $\mu$ F 0.35 % of reading + 7.8 $\mu$ F 0.58 % of reading + 23 $\mu$ F 0.85 % of reading + 77 $\mu$ F	Fluke Multiproduct Calibrator
Capacitance – Measure <sup>1</sup>	1 pF to 6.4 nF (6.4 to 100) nF 100 nF to 1.6 $\mu$ F (1.6 to 100) $\mu$ F	1.2 mF/F 3.8 mF/F 8.1 mF/F 5.7 mF/F	General Radio 1689M RLC Bridge
Inductance – Source/Measure <sup>1</sup> (100 Hz – 1 kHz)	1 mH to 11.11 H	2.7 mH/H	General Radio 1689M RLC Bridge, General Radio 1490D Decade Inductor
Resistance – Source <sup>1</sup>	(0 to 11) $\Omega$ (11 to 33) $\Omega$ (33 to 110) $\Omega$ (110 to 330) $\Omega$ (0.33 to 1.1) k $\Omega$ (1.1 to 3.3) k $\Omega$ (3.3 to 11) k $\Omega$ (11 to 33) k $\Omega$ (33 to 110) k $\Omega$ (110 to 330) k $\Omega$	33 $\mu\Omega/\Omega$ + 0.78 m $\Omega$ 24 $\mu\Omega/\Omega$ + 1.2 m $\Omega$ 22 $\mu\Omega/\Omega$ + 1.1 m $\Omega$ 22 $\mu\Omega/\Omega$ + 1.6 m $\Omega$ 23 $\mu\Omega/\Omega$ + 1.6 $\mu\Omega$ 23 m $\Omega/\Omega$ + 16 m $\Omega$ 23 m $\Omega/\Omega$ + 16 m $\Omega$ 22 m $\Omega/\Omega$ + 0.16 $\Omega$ 23 m $\Omega/\Omega$ + 0.16 $\Omega$ 26 m $\Omega/\Omega$ + 1.6 $\Omega$	Fluke Multiproduct Calibrator





ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source <sup>1</sup>	(0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ (330 to 1 100) MΩ	25 mΩ/Ω + 1.6 Ω 48 mΩ/Ω + 23 Ω 0.23 Ω/Ω + 39 Ω 0.2 Ω/Ω + 1.9 kΩ 0.4 Ω/Ω + 2.3 kΩ 2.3 Ω/kΩ + 78 kΩ 12 Ω/kΩ + 0.39 MΩ	Fluke Multiproduct Calibrator
Resistance – Measure <sup>1</sup>	(0 to 12) Ω (12 to 120) Ω (0.12 to 1.2) kΩ (1.2 to 12) kΩ (1.2 to 120) kΩ (0.12 to 1.2) MΩ (1.2 to 12) MΩ (12 to 120) MΩ (0.12 to 1.2) GΩ	19 μΩ/ Ω + 50 μΩ 15 μΩ/ Ω + 0.5 mΩ 13 μΩ/ Ω + 0.5 mΩ 12 μΩ/ Ω + 5 mΩ 13 μΩ/ Ω + 50 μΩ 24 μΩ/ Ω + 2 Ω 65 μΩ/ Ω + 100 Ω 0.58 mΩ/ Ω + 1 kΩ 5.8 mΩ/ Ω + 10 kΩ	Keysight 3458A 8.5 Digit Multimeter
AC Resistance – Source/Measure <sup>1</sup> (1 kHz)	1 Ω to 100 kΩ	0.39 mΩ/Ω	General Radio 1689M RLC Bridge, HP 16074A Standard Set
Oscilloscopes <sup>1,2</sup>			
DC Voltage into 50 Ω into 1 M Ω	± (0 to 6.6) V ± (0 to 130) V	0.25 % of reading + 40 μV 0.05 % of reading + 40 μV	Fluke Multiproduct Calibrator
Square Wave into 50 Ω into 1 M Ω	± 1 mV to 6.6 V p-p ± 1 mV to 130 V p-p	0.25 % of reading + 40 μV 0.1 % of reading + 40 μV	
Time Markers into 50 Ω	1 ns to 20 ms 50 ms to 5 s	2.5 μs/s (25 + 1 000t) μs/s	



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes <sup>1,2</sup> Leveled Sine Wave Frequency	50 kHz to 1.1 GHz	0.25 $\mu$ Hz/Hz	Fluke Multiproduct Calibrator
Flatness 50 kHz (reference)	5 mV to 5.5 V	2 % of reading + 0.3 mV	
50 kHz to 100 MHz	5 mV to 5.5 V	3.5 % of reading + 0.3 mV	
(100 to 300) MHz	5 mV to 5.5 V	4 % of reading + 0.3 mV	
(300 to 600) MHz	5 mV to 5.5 V	6 % of reading + 0.3 mV	
(0.6 to 1.1) GHz	5 mV to 3.5 V	7 % of reading + 0.3 mV	
Edge Characteristics into 50 $\Omega$ Rise Time	$\leq 300$ ps	+ 0 ps / - 100 ps	
Amplitude	5 mV to 2.5 V	2 % of reading + 0.2 mV	
Frequency	900 Hz to 11 MHz	2.5 $\mu$ Hz/Hz	
Wave Generator Amplitude (Square, Sine, Triangle) into 50 $\Omega$	1.8 mV to 2.5 Vp-p	30 mV/V + 0.1 mV	
into 1 M $\Omega$	1.8 mV to 55 Vp-p	30 mV/V + 0.1 mV	
Frequency	10 Hz to 100 kHz	25 $\mu$ Hz/Hz + 15 mHz	
Input Resistance Measurement into 50 $\Omega$	(40 to 60) $\Omega$	0.1 % of reading	
into 1 M $\Omega$	500 k $\Omega$ to 1.5 M $\Omega$	0.1 % of reading	
Input Capacitance Measurement into 1 M $\Omega$	(5 to 50) pF	5 % of reading + 0.5 pF	



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type B		Fluke Multiproduct Calibrator
	(600 to 800) °C	0.34 °C	
	(800 to 1 000) °C	0.27 °C	
	(1 000 to 1 550) °C	0.23 °C	
	(1 550 to 1 820) °C	0.26 °C	
	Type C		
	(0 to 150) °C	0.23 °C	
	(150 to 650) °C	0.2 °C	
	(650 to 1 000) °C	0.24 °C	
	(1 000 to 1 800) °C	0.39 °C	
	(1 800 to 2 316) °C	0.65 °C	
	Type E		
	(-250 to -100) °C	0.39 °C	
	(-100 to -25) °C	0.13 °C	
	(-25 to 350) °C	0.11 °C	
	(350 to 650) °C	0.13 °C	
	(650 to 1 000) °C	0.17 °C	
	Type J		
	(-210 to -100) °C	0.21 °C	
	(-100 to -30) °C	0.13 °C	
	(-30 to 150) °C	0.11 °C	
(150 to 760) °C	0.14 °C		
(760 to 1 200) °C	0.18 °C		
Type K			
(-200 to -100) °C	0.26 °C		
(-100 to -25) °C	0.14 °C		
(-25 to 120) °C	0.13 °C		
(120 to 1 000) °C	0.2 °C		
(1 000 to 1 372) °C	0.31 °C		
Type L			
(-200 to -100) °C	0.29 °C		
(-100 to 800) °C	0.2 °C		
(800 to 900) °C	0.14 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type N		Fluke Multiproduct Calibrator
	(-200 to -100) °C	0.31 °C	
	(-100 to -25) °C	0.17 °C	
	(-25 to 120) °C	0.15 °C	
	(120 to 410) °C	0.14 °C	
	(410 to 1 300) °C	0.21 °C	
	Type R		
	(0 to 250) °C	0.44 °C	
	(250 to 400) °C	0.27 °C	
	(400 to 1 000) °C	0.26 °C	
	(1 000 to 1 767) °C	0.31 °C	
	Type S		
	(0 to 250) °C	0.37 °C	
	(250 to 1 000) °C	0.28 °C	
	(1 000 to 1 400) °C	0.29 °C	
	(1 400 to 1 767) °C	0.36 °C	
	Type T		
(-250 to -150) °C	0.49 °C		
(-150 to 0) °C	0.19 °C		
(0 to 120) °C	0.13 °C		
(120 to 400) °C	0.11 °C		
Type U			
(-200 to 0) °C	0.44 °C		
(0 to 600) °C	0.21 °C		
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Pt 385, 100 Ω		Fluke Multiproduct Calibrator
	(-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.04 °C	
	(0 to 100) °C	0.05 °C	
	(100 to 300) °C	0.07 °C	
	(300 to 400) °C	0.08 °C	
	(400 to 630) °C	0.09 °C	
	(630 to 800) °C	0.18 °C	
	Pt 3926, 100 Ω		
	(-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.04 °C	
	(0 to 100) °C	0.05 °C	
	(100 to 300) °C	0.07 °C	
	(300 to 400) °C	0.08 °C	
	(400 to 630) °C	0.09 °C	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Pt 3916, 100 Ω		Fluke Multiproduct Calibrator
	(-200 to -190) °C	0.19 °C	
	(-190 to -80) °C	0.03 °C	
	(-80 to 0) °C	0.04 °C	
	(0 to 100) °C	0.05 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.06 °C	
	(300 to 400) °C	0.07 °C	
	(400 to 600) °C	0.08 °C	
	(600 to 630) °C	0.18 °C	
	Pt 385, 200 Ω		
	(-200 to -80) °C	0.03 °C	
	(-80 to 0) °C	0.03 °C	
	(0 to 100) °C	0.03 °C	
	(100 to 260) °C	0.04 °C	
	(260 to 300) °C	0.09 °C	
	(300 to 400) °C	0.1 °C	
	(400 to 600) °C	0.11 °C	
	(600 to 630) °C	0.12 °C	
	Pt 385, 500 Ω		
	(-200 to -80) °C	0.03 °C	
	(-80 to 0) °C	0.04 °C	
	(0 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.06 °C	
	(300 to 400) °C	0.06 °C	
	(400 to 600) °C	0.07 °C	
	(600 to 630) °C	0.09 °C	
	Pt 385, 1 000 Ω		
	(-200 to -80) °C	0.02 °C	
(-80 to 0) °C	0.02 °C		
(0 to 100) °C	0.03 °C		
(100 to 260) °C	0.04 °C		
(260 to 300) °C	0.05 °C		
(300 to 400) °C	0.05 °C		
(400 to 600) °C	0.05 °C		
(600 to 630) °C	0.18 °C		
Ni 385, 120 Ω			
(-80 to 0) °C	0.06 °C		
(0 to 100) °C	0.06 °C		
(100 to 260) °C	0.01 °C		



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Cu 427, 10 Ω (-100 to 260) °C	0.31 °C	Fluke Multiproduct Calibrator

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF/Microwave Power – Source/Measure <sup>1</sup>	100 kHz to 4.2 GHz (-30 to +20) dBm	0.13 dB	Keysight 8482A Power Sensor, Keysight E4418B Power Meter, Signal Generator
	10 MHz to 26.5 GHz (-30 to +20) dBm	0.11 dB	Keysight 8485A Power Sensor, Keysight E4418B Power Meter, Signal Generator
	10 MHz to 18 GHz (-70 to -20) dBm	0.13 dB	Keysight 8481D Power Sensor, Keysight E4418B Power Meter, Signal Generator
Attenuation – Measure or Tuned RF Power 2.5 MHz to 18 GHz	(0 to 10) dB	0.15 dB	Keysight 8902A Measuring Receiver, Keysight 11793A Microwave Converter
	(-10 to 0) dB	0.15 dB	
	(-20 to -10) dB	0.15 dB	
	(-30 to -20) dB	0.15 dB	
	(-40 to -30) dB	0.15 dB	
	(-50 to -40) dB	0.15 dB	
	(-60 to -50) dB	0.17 dB	
	(-70 to -60) dB	0.17 dB	
	(-80 to -70) dB	0.2 dB	
	(-90 to -80) dB	0.22 dB	
	(-100 to -90) dB	0.22 dB	
	(-110 to -100) dB	0.33 dB	
	(-120 to -110) dB	0.43 dB	
(-127 to -120) dB	0.43 dB		

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Amplitude Modulation – Measure 150 kHz to 10 MHz 10 MHz to 1.3 GHz 10 MHz to 1.3 GHz (1.3 to 26.5) GHz (1.3 to 26.5) GHz	Rate: 50 Hz to 10 kHz Depths: (5 to 99) % Rate: 50 Hz to 50 kHz Depths: (5 to 99) % Rate: 20 Hz to 100 kHz Depths: (5 to 99) % Rate: 20 Hz to 100 kHz Depths: (5 to 99) % Rate: 20 Hz to 100 kHz Depths: (5 to 99) %	2.3 % of reading 1.2 % of reading 3.5 % of reading 1.8 % of reading 3.5 % of reading	Agilent 8902A Measuring Receiver
Frequency Modulation – Measure Carrier Frequency: 250 kHz to 10 MHz 10 MHz to 1.3 GHz	Rate: 20 Hz to 10 kHz Dev.: ≤ 40 kHz peak Rate: 20 Hz to 200 kHz Dev.: ≤ 400 kHz peak	2.3 % of reading 5.8 % of reading	Keysight 8902A Measuring Receiver
Frequency Modulation – Measure Carrier Frequency: (1.3 to 26.5) GHz	Rate: 20 Hz to 200 kHz Dev.: ≤ 400 kHz peak	5.8 % of reading	Keysight 8902A Measuring Receiver
Phase Modulation – Measure 10 MHz to 1.3 GHz 1.3 GHz to 26.5 GHz	Rate: 200 Hz 20 kHz Dev.: .1 to 400 rad Rate: 200 Hz 20 kHz Dev.: > .1 to 400 rad	3.5 % of reading 4.6 % of reading	Agilent 8902A Measuring Receiver, Keysight 11722A / 11792A Sensor Modules

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Height Gage <sup>1,2</sup>	Up to 24 in	(22 + 13L) μin	Gage Blocks, Height Gage, Repeat-O-Meter
Micrometer <sup>1</sup>	Up to 1 in	16 μin/in	Gage Blocks
Calipers <sup>1,2</sup>	Up to 36 in	(345 + 5L) μin	Gage Blocks



**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Caliper – Parallelism <sup>1</sup>	(0.15 to 0.5) in	8.8 μin/in	Pin Gage
External Diameter <sup>2</sup> (Pin Gages, Set Plugs)	Up to 12 in	(4.5 + 5.6L) μin	Universal Length Measuring Machine
Internal Diameter <sup>2</sup>	(0.04 to 3) in	(9.4 + 3.2L) μin	Universal Length Measuring Machine
Radius Gages <sup>2</sup>	(0.031 25 to 1) in	(291 + 98L) μin	Keyence
Gage Blocks <sup>2</sup>	Up to 12 in	(1.5 + 5.7L) μin	Universal Length Measuring Machine, Master Gage Blocks
Angle Measurement <sup>2</sup> Digital Protractors, Levels, Inclinometers	Up to 90°	22" (0.006°)	Sine Bar, Gage Blocks, Surface Plate, Right Angles

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Class I Balances <sup>1</sup> (0.000 01 g resolution)	Up to 100 g	0.29 mg	ASTM E617 Class 1 Weights and NIST Handbook 44 utilized for the calibration of the weighing system.
(0.000 02 g resolution)	Up to 100 g	0.29 mg	
(0.000 05 g resolution)	Up to 100 g	0.3 mg	
(0.000 1 g resolution)	Up to 200 g	0.6 mg	
(0.000 2 g resolution)	Up to 200 g	0.6 mg	
(0.000 5 g resolution)	Up to 200 g	0.87 mg	

**Mass and Mass Related**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Class II Balances <sup>1</sup> (0.001 g resolution) (0.002 g resolution) (0.005 g resolution) (0.01 g resolution) (0.02 g resolution) (0.05 g resolution) (0.1 g resolution) (0.2 g resolution)	Up to 100 g Up to 200 g Up to 500 g Up to 1 kg Up to 2 kg Up to 5 kg Up to 10 kg Up to 20 kg	1.3 mg 2.6 mg 6.6 mg 13 mg 29 mg 0.15 g 0.31 g 0.36 g	ASTM E617 Class 1 Weights and NIST Handbook 44 utilized for the calibration of the weighing system.
Class III Light Capacity Scales <sup>1</sup> (0.000 5 lb resolution) (0.001 lb resolution) (0.002 lb resolution) (0.005 lb resolution) (0.01 lb resolution) (0.02 lb resolution) (0.05 lb resolution) (0.1 lb resolution)	Up to 5 lb Up to 10 lb Up to 20 lb Up to 50 lb Up to 100 lb Up to 200 lb Up to 500 lb Up to 500 lb	0.000 87 lb 0.001 7 lb 0.002 6 lb 0.008 7 lb 0.018 lb 0.036 lb 0.087 lb 0.14 lb	NIST Class F Weights and NIST Handbook 44 utilized for the calibration of the weighing system.
Force Gages – Tension <sup>1</sup>	(1 to 110) lbf	0.06 lb	NIST Class F Weights, Hangers
Durometers – Spring Force Only Types A, B, O, C, D, DO	209.04 gf to 4.079 kgf	0.01 % of reading	ASTM E617 Class 1 Weights, Triple Beam Balance
Pressure – Source/Measure <sup>1</sup>	(-5 to 5) psi	0.081 psi	Comparison to Fluke 700PD3 Pressure Module

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure – Source/Measure <sup>1</sup>	(-12 to 100) psi	0.19 psi	Fluke 718-300 Pressure Calibrator
	Up to 1 000 psig	0.66 psi	Comparison to Crystal XP2i Master Digital Pressure Gage
	Up to 3 000 psig	2.4 psi	Comparison to Druck DPI 104-3000 Digital Pressure Gage
	Up to 10 000 psig	6 psi	Comparison to Druck DPI 104-10000 Digital Pressure Gage
	Up to 10 inH <sub>2</sub> O	0.3 inH <sub>2</sub> O	Comparison to Fluke 700PO1 Pressure Module
Torque Indicating Device <sup>1</sup>	(40 to 200) ozf·in (4 to 50) lbf·in (30 to 400) lbf·in (80 to 1 000) lbf·in (20 to 250) lbf·ft (60 to 600) lbf·ft	0.29 % of reading 0.29 % of reading 0.29 % of reading 0.3 % of reading 0.34 % of reading 0.59 % of reading	Torque Transducers
Torque Analyzers, Torque Transducers	(40 to 200) ozf·in (4 to 50) lbf·in (30 to 400) lbf·in (80 to 1 000) lbf·in (20 to 250) lbf·ft (60 to 600) lbf·ft	0.044 % of reading 0.044 % of reading 0.044 % of reading 0.044 % of reading 0.054 % of reading 0.096 % of reading	Torque Wheels/Arms, Master Weights, Weight Hanger

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Generate/Measure <sup>1</sup>	(-25 to 420) °C	0.04 °C	Fluke Blackstack Readout, SPRT, Heat Source
	(400 to 1 200) °C	0.32 °C	Fluke Blackstack Readout, Type-S Thermocouple Probe, Heat Source

### Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Infrared Thermometers <sup>1</sup>	100 °C 300 °C 500 °C	0.6 °C 0.85 °C 1.2 °C	Comparison to SPRT w/ Fluke 9132A Blackbody Calibrator (flat plate) $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$
Humidity – Generate	(10 to 90) %RH	1.2 %RH	Thunder Scientific, Psychrometer

### Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source/Measure	10 Hz to 26.5 GHz	50 pHz/Hz	GPS Disciplined Oscillator, Signal Generators, Universal Counters

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2.  $L$  = length in inches,  $T$  = time in seconds; " = arc-second.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2548.



R. Douglas Leonard Jr., VP, PILR SBU